

Diversey Europe Operations EV Maarssenbroeksed!jk 2 3542 DN Utrecht The Netherlands

Tel. +31 (0)30 247 6427 www.Diversey.com

CERTIFICATE OF ANALYSIS

Date:

26 July 2022

Product Name	CLEARKLENS BI-SPOR	CLEARKLENS BI-SPORE ACTIVATOR SOLUTION VH26 first off		
Product Code	7515828			
Batch Number	FMP22199	54445		
Production Date	18/07/2022			
Expiration Date	EXP 18/07/2024			

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limi Lower		Results
Appearance	Visual	Clear Colou	rless Liquid	Clear Colourless Liquid
pH (neat solution)	DM001	9.0	12.5	11.5
Specific Gravity (20°C)	DM004	1.004	1.020	1.005
Potential Chlorine Dioxide		100	180	138.375

On behalf of Diversey site Quality Manger	Name:	urszula Haraburda Angelika Partynska
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA Template Version: 02 Date of issuing: November 24th 2017



Diversey Europe Operations EV Maarssenbroeksedijk 2 3542 DN Utrecht The Netherlands

Tel. +31 (0)30 247 6427 www.Diversey.com

CERTIFICATE OF ANALYSIS

Date:

26 July 2022

Product Name	CLEARKLENS BI-SPORE BASE SOLUTION VH26 first off 7515828		
Product Code			
Batch Number	FMP22199	54445	
Production Date	18/07/2022		
Expiration Date	EXP 18/07/2024		

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limi Lower	ts - Upper	Results
Appearance	Visual	Clear Colou	rless Liquid	Clear Colourless Liquid
pH (neat solution)	DM001	1.5	2.5	2.0
Specific Gravity (20°C)	DM004	1.010	1.030	1.015

On behalf of Diversey site Quality Manger	Name:	urszula Haraburda Angelika Partynska
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA Template Version: 02 Date of issuing: November 24th 2017

STERIS: Gamma Certificate Of Processing

Prepared for: FLEXIBLE MEDICAL PACKAGING LTD (8245)

Gamma Process Run ID 2173-28061A

Product CodeLot NumberQuantityUOMBI-SP B3 Components DV5857545068Case

Validation Reference Number: 5857

Specified Dose Range (kGy): 25.0 - 40.0 Calculated Min Dose (kGy): 30.4 Reference Dose Range (kGy): 26.1 - 35.8 Calculated Max Dose (kGy): 36.7

PO Number: 402495

Product meets Customer specifications; zero nonconformities occurred during this irradiation run.

Signature Manifest

Reviewed and E-Signed By: Tamas Szatmari (Quality Engineer)

Date/Time E-Signed: 09-May-2022 12:48 PM

Document Content Revision: 2

Operating facilities are in compliance with applicable regulations providing services under a certified quality system which meets the requirements of FDA QSR, EN/ISO 13485 current certified version, and is in alignment with EN ISO 11137 current certified version

Processing Location

Synergy Health Sterilisation UK Limited, a STERIS Company Brunel Close Drayton Fields Industrial Estate Daventry Northants NN11 8RB

Phone: + 44(0) 1327 706 111

Items irradiated under WO54506 will be used in finished batch Fmp22199 54445.

Olga Kirchner

Document ID: 111139

N/A Last Revised in Rel 2.0.0.0 Rel Date: 13-Aug-2018 Page 1 of 1

PHARMACEUTICAL ANALYSIS REPORT



Flexible Medical Packaging

Unit 8 Hightown White Cross Industrial Estate Lancaster LA14XS

FAO:

Mrs.Olga Kirchner

Report of Tests on:

Bi-Spore Activator

Sample Description:

Sample Code: Fmp22199 54445

Lucideon Sample Number: UK223076-19812

Lucideon Report Number: UK223076-19812/MFEP

Issue Number:

Date Logged:

29-Jul-2022

Order Number:

403116

Date Reported:

19-Aug-2022

Date(s) of Test(s):

04-Aug-2022 to 19-Aug-2022

Sterility Testing Membrane Filtration EP

Test Results:

Result: Pass

The test results meet the EP/USP criteria: Yes

Tests carried out in accordance with cGMP Refer to Deviation (DEV) 6557.

End of Test Report

BCOK 19-Rg-22

Mrs Natalie Boot

Senior Business Support Administrator

This report is issued in accordance with the Conditions of Business of Lucideon Limited and relates only to the sample(s) tested. No responsibility is taken for the accuracy of the sampling unless this is done under our own supervision. This report shall not be

PHARMACEUTICAL ANALYSIS REPORT



Flexible Medical Packaging

Unit 8 Hightown White Cross Industrial Estate Lancaster LA14XS

FAO:

Mrs.Olga Kirchner

Report of Tests on:

Bi-Spore Base

Sample Description:

Sample Code: Fmp22199 54445

Lucideon Sample Number: UK223076-19813

Lucideon Report Number: UK223076-19813/MFEP

Issue Number:

1

Date Logged:

29-Jul-2022

Order Number:

403116

Date Reported:

19-Aug-2022

Date(s) of Test(s):

04-Aug-2022 to 19-Aug-2022

Sterility Testing

Membrane Filtration EP

Test Results:

Result: Pass

The test results meet the EP/USP criteria: Yes

Tests carried out in accordance with cGMP Refer to Deviation (DEV) 6557.

End of Test Report

19-Ag-22

Mrs Natalie Boot

Senior Business Support Administrator